

Teledyne FLIR LLC - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: MF100
Description: AC/DC Magnetic Field Meter
Date of Issue: 29-Sep-21

We, Teledyne FLIR LLC - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:


EMC Directive: 2014/30/EU
Report Number: 110117-O.EMI
Report Date of Issue: 1/17/2011

Standards:

EN 55022 Class B
EN 61000-4-2
EN 61000-4-3

RoHS Directive: EU Directive 2015/863/EU (RoHS 3)

The test reports show that the product fulfills the requirement in the EC EMC Directive and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC EMC Directive and RoHS Directive.



Mark Sultzbach / QA Manager